



February 18, 2025

Terry Stoddard
Citi Boces
179 County Route 64
Mexico, NY 13114

RE: Project: CITI BOCES 2/6
Pace Project No.: 70337400

Dear Terry Stoddard:

Enclosed are the analytical results for sample(s) received by the laboratory on February 11, 2025. The results relate only to the samples included in this report. Results reported herein conform to the applicable TNI/NELAC Standards and the laboratory's Quality Manual, where applicable, unless otherwise noted in the body of the report.

The test results provided in this final report were generated by each of the following laboratories within the Pace Network:

- Pace Analytical Services - Melville

If you have any questions concerning this report, please feel free to contact me.

Sincerely,

Alexandria Correa for
Meghan Heister
meghan.heister@pacelabs.com
516-370-6000
Project Manager

Enclosures

cc: Ian Rowberry, Citi Boces
Citi BOCES Safety and, Citi Boces_The Center for
instruction, Tech. and Innovation



REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



CERTIFICATIONS

Project: CITI BOCES 2/6

Pace Project No.: 70337400

Pace Analytical Services, LLC - Melville, NY

575 Broad Hollow Rd, Melville, NY 11747

Connecticut Certification #: PH-0435

Delaware Certification # NY 10478

Maryland Certification #: 208

Massachusetts Certification #: M-NY026

New Hampshire Certification #: 2987

New Jersey Certification #: NY158

New York Certification #: 10478 Primary Accrediting Body

Pennsylvania Certification #: 68-00350

Rhode Island Certification #: LAO00340

Texas Certification #: T104704582

Florida Certification #: E871198

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: CITI BOCES 2/6

Pace Project No.: 70337400

Sample: CITI 24		Lab ID: 70337400001		Collected: 02/06/25 05:54		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.1	ug/L	1.0	1		02/13/25 12:50	7439-92-1		
Sample: CITI 27		Lab ID: 70337400002		Collected: 02/06/25 06:11		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		02/13/25 12:52	7439-92-1		
Sample: CITI 28		Lab ID: 70337400003		Collected: 02/06/25 06:11		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		02/13/25 12:53	7439-92-1		
Sample: CITI 35		Lab ID: 70337400004		Collected: 02/06/25 06:05		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		02/13/25 12:55	7439-92-1		
Sample: CITI 36		Lab ID: 70337400005		Collected: 02/06/25 06:03		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.6	ug/L	1.0	1		02/13/25 12:59	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: CITI BOCES 2/6

Pace Project No.: 70337400

Sample: CITI 37		Lab ID: 70337400006	Collected: 02/06/25 06:03	Received: 02/11/25 06:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	2.0	ug/L	1.0	1		02/13/25 13:00	7439-92-1	
------	-----	------	-----	---	--	----------------	-----------	--

Sample: CITI 38		Lab ID: 70337400007	Collected: 02/06/25 06:03	Received: 02/11/25 06:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	1.3	ug/L	1.0	1		02/13/25 13:02	7439-92-1	
------	-----	------	-----	---	--	----------------	-----------	--

Sample: CITI 40		Lab ID: 70337400008	Collected: 02/06/25 06:05	Received: 02/11/25 06:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 13:03	7439-92-1	
------	------	------	-----	---	--	----------------	-----------	--

Sample: CITI 43		Lab ID: 70337400009	Collected: 02/06/25 06:14	Received: 02/11/25 06:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 13:05	7439-92-1	
------	------	------	-----	---	--	----------------	-----------	--

Sample: CITI 44		Lab ID: 70337400010	Collected: 02/06/25 06:14	Received: 02/11/25 06:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 13:06	7439-92-1	
------	------	------	-----	---	--	----------------	-----------	--

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: CITI BOCES 2/6

Pace Project No.: 70337400

Sample: CITI 45		Lab ID: 70337400011	Collected: 02/06/25 06:14		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 13:08	7439-92-1	
------	------	------	-----	---	--	----------------	-----------	--

Sample: CITI 48		Lab ID: 70337400012	Collected: 02/06/25 06:39		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 13:09	7439-92-1	
------	------	------	-----	---	--	----------------	-----------	--

Sample: CITI 52		Lab ID: 70337400013	Collected: 02/06/25 06:41		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 13:11	7439-92-1	
------	------	------	-----	---	--	----------------	-----------	--

Sample: CITI 57		Lab ID: 70337400014	Collected: 02/06/25 06:44		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 13:23	7439-92-1	
------	------	------	-----	---	--	----------------	-----------	--

Sample: CITI 58		Lab ID: 70337400015	Collected: 02/06/25 06:42		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 13:27	7439-92-1	
------	------	------	-----	---	--	----------------	-----------	--

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: CITI BOCES 2/6

Pace Project No.: 70337400

Sample: CITI 60		Lab ID: 70337400016		Collected: 02/06/25 06:06		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.0	ug/L	1.0	1		02/13/25 13:31	7439-92-1		
Sample: CITI 61		Lab ID: 70337400017		Collected: 02/06/25 06:47		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		02/13/25 13:33	7439-92-1		
Sample: CITI 64		Lab ID: 70337400018		Collected: 02/06/25 06:53		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		02/13/25 13:34	7439-92-1		
Sample: CITI 65		Lab ID: 70337400019		Collected: 02/06/25 06:53		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		02/13/25 13:39	7439-92-1		
Sample: CITI 68		Lab ID: 70337400020		Collected: 02/06/25 06:48		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	39.1	ug/L	1.0	1		02/13/25 13:40	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: CITI BOCES 2/6

Pace Project No.: 70337400

Sample: CITI 69		Lab ID: 70337400021	Collected: 02/06/25 06:50		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	1.3	ug/L	1.0	1		02/13/25 13:42	7439-92-1	
------	-----	------	-----	---	--	----------------	-----------	--

Sample: CITI 70		Lab ID: 70337400022	Collected: 02/06/25 06:51		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	1.0	ug/L	1.0	1		02/13/25 13:43	7439-92-1	
------	-----	------	-----	---	--	----------------	-----------	--

Sample: CITI 71		Lab ID: 70337400023	Collected: 02/06/25 06:35		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	1.7	ug/L	1.0	1		02/13/25 13:44	7439-92-1	
------	-----	------	-----	---	--	----------------	-----------	--

Sample: CITI 72		Lab ID: 70337400024	Collected: 02/06/25 06:35		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 13:46	7439-92-1	
------	------	------	-----	---	--	----------------	-----------	--

Sample: CITI 73		Lab ID: 70337400025	Collected: 02/06/25 06:36		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	1.2	ug/L	1.0	1		02/13/25 13:47	7439-92-1	
------	-----	------	-----	---	--	----------------	-----------	--

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: CITI BOCES 2/6

Pace Project No.: 70337400

Sample: CITI 74		Lab ID: 70337400026	Collected: 02/06/25 06:37		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 13:49	7439-92-1	
------	------	------	-----	---	--	----------------	-----------	--

Sample: CITI 79		Lab ID: 70337400027	Collected: 02/06/25 06:22		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 13:50	7439-92-1	
------	------	------	-----	---	--	----------------	-----------	--

Sample: CITI 80		Lab ID: 70337400028	Collected: 02/06/25 06:22		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 13:52	7439-92-1	
------	------	------	-----	---	--	----------------	-----------	--

Sample: CITI 87		Lab ID: 70337400029	Collected: 02/06/25 06:25		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 13:56	7439-92-1	
------	------	------	-----	---	--	----------------	-----------	--

Sample: CITI 88		Lab ID: 70337400030	Collected: 02/06/25 06:25		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 13:58	7439-92-1	
------	------	------	-----	---	--	----------------	-----------	--

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: CITI BOCES 2/6

Pace Project No.: 70337400

Sample: CITI 89		Lab ID: 70337400031	Collected: 02/06/25 06:25		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 13:59	7439-92-1	
------	------	------	-----	---	--	----------------	-----------	--

Sample: CITI 90		Lab ID: 70337400032	Collected: 02/06/25 06:26		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	18.9	ug/L	1.0	1		02/13/25 14:00	7439-92-1	
------	------	------	-----	---	--	----------------	-----------	--

Sample: CITI 93		Lab ID: 70337400033	Collected: 02/06/25 05:24		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 14:02	7439-92-1	
------	------	------	-----	---	--	----------------	-----------	--

Sample: CITI 96		Lab ID: 70337400034	Collected: 02/06/25 05:25		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 14:06	7439-92-1	
------	------	------	-----	---	--	----------------	-----------	--

Sample: CITI 97		Lab ID: 70337400035	Collected: 02/06/25 05:25		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 14:14	7439-92-1	
------	------	------	-----	---	--	----------------	-----------	--

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: CITI BOCES 2/6

Pace Project No.: 70337400

Sample: CITI 98		Lab ID: 70337400036		Collected: 02/06/25 05:26		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 14:18	7439-92-1
------	------	------	-----	---	--	----------------	-----------

Sample: CITI 99		Lab ID: 70337400037		Collected: 02/06/25 05:26		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 14:19	7439-92-1
------	------	------	-----	---	--	----------------	-----------

Sample: CITI 100		Lab ID: 70337400038		Collected: 02/06/25 05:27		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 14:21	7439-92-1
------	------	------	-----	---	--	----------------	-----------

Sample: CITI 101		Lab ID: 70337400039		Collected: 02/06/25 05:28		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 14:22	7439-92-1
------	------	------	-----	---	--	----------------	-----------

Sample: CITI 102		Lab ID: 70337400040		Collected: 02/06/25 05:28		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 14:24	7439-92-1
------	------	------	-----	---	--	----------------	-----------

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: CITI BOCES 2/6

Pace Project No.: 70337400

Sample: CITI 103		Lab ID: 70337400041	Collected: 02/06/25 05:29		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 14:25	7439-92-1	
------	------	------	-----	---	--	----------------	-----------	--

Sample: CITI 104		Lab ID: 70337400042	Collected: 02/06/25 05:29		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 14:27	7439-92-1	
------	------	------	-----	---	--	----------------	-----------	--

Sample: CITI 105		Lab ID: 70337400043	Collected: 02/06/25 05:32		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	3.5	ug/L	1.0	1		02/13/25 14:31	7439-92-1	
------	-----	------	-----	---	--	----------------	-----------	--

Sample: CITI 110		Lab ID: 70337400044	Collected: 02/06/25 05:30		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 14:32	7439-92-1	
------	------	------	-----	---	--	----------------	-----------	--

Sample: CITI 111		Lab ID: 70337400045	Collected: 02/06/25 05:31		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 14:34	7439-92-1	
------	------	------	-----	---	--	----------------	-----------	--

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: CITI BOCES 2/6

Pace Project No.: 70337400

Sample: CITI 112		Lab ID: 70337400046		Collected: 02/06/25 05:29		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 14:35	7439-92-1
------	------	------	-----	---	--	----------------	-----------

Sample: CITI 118		Lab ID: 70337400047		Collected: 02/06/25 05:35		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 14:37	7439-92-1
------	------	------	-----	---	--	----------------	-----------

Sample: CITI 120		Lab ID: 70337400048		Collected: 02/06/25 05:49		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 14:38	7439-92-1
------	------	------	-----	---	--	----------------	-----------

Sample: CITI 122		Lab ID: 70337400049		Collected: 02/06/25 05:41		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 14:40	7439-92-1
------	------	------	-----	---	--	----------------	-----------

Sample: CITI 123		Lab ID: 70337400050		Collected: 02/06/25 05:41		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 14:41	7439-92-1
------	------	------	-----	---	--	----------------	-----------

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: CITI BOCES 2/6

Pace Project No.: 70337400

Sample: CITI 124		Lab ID: 70337400051		Collected: 02/06/25 05:42		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		02/13/25 14:43	7439-92-1		
Sample: CITI 125		Lab ID: 70337400052		Collected: 02/06/25 05:45		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		02/13/25 14:44	7439-92-1		
Sample: CITI 126		Lab ID: 70337400053		Collected: 02/06/25 05:45		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	3.7	ug/L	1.0	1		02/13/25 14:49	7439-92-1		
Sample: CITI 130		Lab ID: 70337400054		Collected: 02/06/25 07:05		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		02/13/25 14:53	7439-92-1		
Sample: CITI 132		Lab ID: 70337400055		Collected: 02/06/25 05:09		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.3	ug/L	1.0	1		02/13/25 14:57	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: CITI BOCES 2/6

Pace Project No.: 70337400

Sample: CITI 133		Lab ID: 70337400056		Collected: 02/06/25 05:05		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	1.4	ug/L	1.0	1		02/13/25 15:02	7439-92-1		
------	-----	------	-----	---	--	----------------	-----------	--	--

Sample: CITI 137		Lab ID: 70337400057		Collected: 02/06/25 07:06		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 15:06	7439-92-1		
------	------	------	-----	---	--	----------------	-----------	--	--

Sample: CITI 142		Lab ID: 70337400058		Collected: 02/06/25 07:12		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 15:07	7439-92-1		
------	------	------	-----	---	--	----------------	-----------	--	--

Sample: CITI 145		Lab ID: 70337400059		Collected: 02/06/25 05:04		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	5.8	ug/L	1.0	1		02/13/25 15:09	7439-92-1		
------	-----	------	-----	---	--	----------------	-----------	--	--

Sample: CITI 149		Lab ID: 70337400060		Collected: 02/06/25 07:21		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	2.3	ug/L	1.0	1		02/13/25 15:10	7439-92-1		
------	-----	------	-----	---	--	----------------	-----------	--	--

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: CITI BOCES 2/6

Pace Project No.: 70337400

Sample: CITI 152		Lab ID: 70337400061	Collected: 02/06/25 07:23		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 15:12	7439-92-1	
------	------	------	-----	---	--	----------------	-----------	--

Sample: CITI 153		Lab ID: 70337400062	Collected: 02/06/25 07:25		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 15:13	7439-92-1	
------	------	------	-----	---	--	----------------	-----------	--

Sample: CITI 157		Lab ID: 70337400063	Collected: 02/06/25 07:27		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	1.5	ug/L	1.0	1		02/13/25 15:15	7439-92-1	
------	-----	------	-----	---	--	----------------	-----------	--

Sample: CITI 160		Lab ID: 70337400064	Collected: 02/06/25 07:31		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	<1.0	ug/L	1.0	1		02/13/25 15:16	7439-92-1	
------	------	------	-----	---	--	----------------	-----------	--

Sample: CITI 163		Lab ID: 70337400065	Collected: 02/06/25 06:57		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS Drinking Water
Analytical Method: EPA 200.8
Pace Analytical Services - Melville

Lead	1.9	ug/L	1.0	1		02/13/25 15:18	7439-92-1	
------	-----	------	-----	---	--	----------------	-----------	--

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: CITI BOCES 2/6
Pace Project No.: 70337400

Sample: CITI 173		Lab ID: 70337400066		Collected: 02/06/25 07:41		Received: 02/11/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8							
		Pace Analytical Services - Melville							
Lead	1.0	ug/L	1.0	1		02/13/25 15:19	7439-92-1		

REPORT OF LABORATORY ANALYSIS



QUALITY CONTROL DATA

Project: CITI BOCES 2/6

Pace Project No.: 70337400

QC Batch:	384690	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70337400001, 70337400002, 70337400003, 70337400004, 70337400005, 70337400006, 70337400007, 70337400008, 70337400009, 70337400010, 70337400011, 70337400012, 70337400013		

METHOD BLANK:	2020528	Matrix:	Water
Associated Lab Samples:	70337400001, 70337400002, 70337400003, 70337400004, 70337400005, 70337400006, 70337400007, 70337400008, 70337400009, 70337400010, 70337400011, 70337400012, 70337400013		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	02/13/25 12:27	

LABORATORY CONTROL SAMPLE:	2020529					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	49.5	99	85-115	

MATRIX SPIKE SAMPLE:		2020532					
		70337368023	Spike	MS	MS	% Rec	
Parameter	Units	Result	Conc.	Result	% Rec	Limits	Qualifiers
Lead	ug/L	<1.0	50	49.5	99	70-130	

MATRIX SPIKE SAMPLE:		2020536					
		70337368024	Spike	MS	MS	% Rec	
Parameter	Units	Result	Conc.	Result	% Rec	Limits	Qualifiers
Lead	ug/L	<1.0	50	49.4	98	70-130	

SAMPLE DUPLICATE:	2020530				
Parameter	Units	70337368023 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

SAMPLE DUPLICATE:	2020534				
Parameter	Units	70337368024 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: CITI BOCES 2/6

Pace Project No.: 70337400

QC Batch:	384720	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70337400014, 70337400015, 70337400016, 70337400017, 70337400018, 70337400019, 70337400020, 70337400021, 70337400022, 70337400023, 70337400024, 70337400025, 70337400026, 70337400027, 70337400028, 70337400029, 70337400030, 70337400031, 70337400032, 70337400033		

METHOD BLANK:	2020684	Matrix:	Water
Associated Lab Samples:	70337400014, 70337400015, 70337400016, 70337400017, 70337400018, 70337400019, 70337400020, 70337400021, 70337400022, 70337400023, 70337400024, 70337400025, 70337400026, 70337400027, 70337400028, 70337400029, 70337400030, 70337400031, 70337400032, 70337400033		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	02/13/25 13:12	

LABORATORY CONTROL SAMPLE:	2020685					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	49.6	99	85-115	

MATRIX SPIKE SAMPLE:		2020687					
		70337400014	Spike	MS	MS	% Rec	
Parameter	Units	Result	Conc.	Result	% Rec	Limits	Qualifiers
Lead	ug/L	<1.0	50	36.2	71	70-130	

MATRIX SPIKE SAMPLE:		2020689					
		70337400015	Spike	MS	MS	% Rec	
Parameter	Units	Result	Conc.	Result	% Rec	Limits	Qualifiers
Lead	ug/L	<1.0	50	47.7	94	70-130	

SAMPLE DUPLICATE:	2020686				
Parameter	Units	70337400014 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

SAMPLE DUPLICATE:	2020688				
Parameter	Units	70337400015 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: CITI BOCES 2/6

Pace Project No.: 70337400

QC Batch:	384721	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70337400034, 70337400035, 70337400036, 70337400037, 70337400038, 70337400039, 70337400040, 70337400041, 70337400042, 70337400043, 70337400044, 70337400045, 70337400046, 70337400047, 70337400048, 70337400049, 70337400050, 70337400051, 70337400052, 70337400053		

METHOD BLANK:	2020692	Matrix:	Water
Associated Lab Samples:	70337400034, 70337400035, 70337400036, 70337400037, 70337400038, 70337400039, 70337400040, 70337400041, 70337400042, 70337400043, 70337400044, 70337400045, 70337400046, 70337400047, 70337400048, 70337400049, 70337400050, 70337400051, 70337400052, 70337400053		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	02/13/25 14:03	

LABORATORY CONTROL SAMPLE:	2020693					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	49.2	98	85-115	

MATRIX SPIKE SAMPLE:		2020695					
		70337400034	Spike	MS	MS	% Rec	
Parameter	Units	Result	Conc.	Result	% Rec	Limits	Qualifiers
Lead	ug/L	<1.0	50	47.7	95	70-130	

MATRIX SPIKE SAMPLE:		2020697					
		70337400035	Spike	MS	MS	% Rec	
Parameter	Units	Result	Conc.	Result	% Rec	Limits	Qualifiers
Lead	ug/L	<1.0	50	49.1	97	70-130	

SAMPLE DUPLICATE:	2020694				
Parameter	Units	70337400034 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

SAMPLE DUPLICATE:	2020696				
Parameter	Units	70337400035 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: CITI BOCES 2/6

Pace Project No.: 70337400

QC Batch:	384724	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70337400054, 70337400055, 70337400056, 70337400057, 70337400058, 70337400059, 70337400060, 70337400061, 70337400062, 70337400063, 70337400064, 70337400065, 70337400066		

METHOD BLANK:	2020705	Matrix:	Water
Associated Lab Samples:	70337400054, 70337400055, 70337400056, 70337400057, 70337400058, 70337400059, 70337400060, 70337400061, 70337400062, 70337400063, 70337400064, 70337400065, 70337400066		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	02/13/25 14:50	

LABORATORY CONTROL SAMPLE:	2020706					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	47.6	95	85-115	

MATRIX SPIKE SAMPLE:		2020708					
		70337400054	Spike	MS	MS	% Rec	
Parameter	Units	Result	Conc.	Result	% Rec	Limits	Qualifiers
Lead	ug/L	<1.0	50	47.4	95	70-130	

MATRIX SPIKE SAMPLE:		2020710					
		70337400055	Spike	MS	MS	% Rec	
Parameter	Units	Result	Conc.	Result	% Rec	Limits	Qualifiers
Lead	ug/L	1.3	50	50.2	98	70-130	

SAMPLE DUPLICATE:	2020707				
Parameter	Units	70337400054 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

SAMPLE DUPLICATE:	2020709				
Parameter	Units	70337400055 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	1.3	1.1	10	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALIFIERS

Project: CITI BOCES 2/6
Pace Project No.: 70337400

DEFINITIONS

DF - Dilution Factor, if reported, represents the factor applied to the reported data due to dilution of the sample aliquot.

ND - Not Detected at or above adjusted reporting limit.

TNTC - Too Numerous To Count

J - Estimated concentration above the adjusted method detection limit and below the adjusted reporting limit.

MDL - Adjusted Method Detection Limit.

PQL - Practical Quantitation Limit.

RL - Reporting Limit - The lowest concentration value that meets project requirements for quantitative data with known precision and bias for a specific analyte in a specific matrix.

S - Surrogate

1,2-Diphenylhydrazine decomposes to and cannot be separated from Azobenzene using Method 8270. The result for each analyte is a combined concentration.

Consistent with EPA guidelines, unrounded data are displayed and have been used to calculate % recovery and RPD values.

LCS(D) - Laboratory Control Sample (Duplicate)

MS(D) - Matrix Spike (Duplicate)

DUP - Sample Duplicate

RPD - Relative Percent Difference

NC - Not Calculable.

SG - Silica Gel - Clean-Up

U - Indicates the compound was analyzed for, but not detected.

N-Nitrosodiphenylamine decomposes and cannot be separated from Diphenylamine using Method 8270. The result reported for each analyte is a combined concentration.

Reported results are not rounded until the final step prior to reporting. Therefore, calculated parameters that are typically reported as "Total" may vary slightly from the sum of the reported component parameters.

Pace Analytical is TNI accredited. Contact your Pace PM for the current list of accredited analytes.

TNI - The NELAC Institute.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA CROSS REFERENCE TABLE

Project: CITI BOCES 2/6

Pace Project No.: 70337400

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
70337400001	CITI 24	EPA 200.8	384690		
70337400002	CITI 27	EPA 200.8	384690		
70337400003	CITI 28	EPA 200.8	384690		
70337400004	CITI 35	EPA 200.8	384690		
70337400005	CITI 36	EPA 200.8	384690		
70337400006	CITI 37	EPA 200.8	384690		
70337400007	CITI 38	EPA 200.8	384690		
70337400008	CITI 40	EPA 200.8	384690		
70337400009	CITI 43	EPA 200.8	384690		
70337400010	CITI 44	EPA 200.8	384690		
70337400011	CITI 45	EPA 200.8	384690		
70337400012	CITI 48	EPA 200.8	384690		
70337400013	CITI 52	EPA 200.8	384690		
70337400014	CITI 57	EPA 200.8	384720		
70337400015	CITI 58	EPA 200.8	384720		
70337400016	CITI 60	EPA 200.8	384720		
70337400017	CITI 61	EPA 200.8	384720		
70337400018	CITI 64	EPA 200.8	384720		
70337400019	CITI 65	EPA 200.8	384720		
70337400020	CITI 68	EPA 200.8	384720		
70337400021	CITI 69	EPA 200.8	384720		
70337400022	CITI 70	EPA 200.8	384720		
70337400023	CITI 71	EPA 200.8	384720		
70337400024	CITI 72	EPA 200.8	384720		
70337400025	CITI 73	EPA 200.8	384720		
70337400026	CITI 74	EPA 200.8	384720		
70337400027	CITI 79	EPA 200.8	384720		
70337400028	CITI 80	EPA 200.8	384720		
70337400029	CITI 87	EPA 200.8	384720		
70337400030	CITI 88	EPA 200.8	384720		
70337400031	CITI 89	EPA 200.8	384720		
70337400032	CITI 90	EPA 200.8	384720		
70337400033	CITI 93	EPA 200.8	384720		
70337400034	CITI 96	EPA 200.8	384721		
70337400035	CITI 97	EPA 200.8	384721		
70337400036	CITI 98	EPA 200.8	384721		
70337400037	CITI 99	EPA 200.8	384721		
70337400038	CITI 100	EPA 200.8	384721		
70337400039	CITI 101	EPA 200.8	384721		
70337400040	CITI 102	EPA 200.8	384721		
70337400041	CITI 103	EPA 200.8	384721		
70337400042	CITI 104	EPA 200.8	384721		
70337400043	CITI 105	EPA 200.8	384721		
70337400044	CITI 110	EPA 200.8	384721		
70337400045	CITI 111	EPA 200.8	384721		
70337400046	CITI 112	EPA 200.8	384721		
70337400047	CITI 118	EPA 200.8	384721		
70337400048	CITI 120	EPA 200.8	384721		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA CROSS REFERENCE TABLE

Project: CITI BOCES 2/6

Pace Project No.: 70337400

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
70337400049	CITI 122	EPA 200.8	384721		
70337400050	CITI 123	EPA 200.8	384721		
70337400051	CITI 124	EPA 200.8	384721		
70337400052	CITI 125	EPA 200.8	384721		
70337400053	CITI 126	EPA 200.8	384721		
70337400054	CITI 130	EPA 200.8	384724		
70337400055	CITI 132	EPA 200.8	384724		
70337400056	CITI 133	EPA 200.8	384724		
70337400057	CITI 137	EPA 200.8	384724		
70337400058	CITI 142	EPA 200.8	384724		
70337400059	CITI 145	EPA 200.8	384724		
70337400060	CITI 149	EPA 200.8	384724		
70337400061	CITI 152	EPA 200.8	384724		
70337400062	CITI 153	EPA 200.8	384724		
70337400063	CITI 157	EPA 200.8	384724		
70337400064	CITI 160	EPA 200.8	384724		
70337400065	CITI 163	EPA 200.8	384724		
70337400066	CITI 173	EPA 200.8	384724		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

Pace® Location Requested (City/State):
Pace Analytical Long Island NY
575 Broad Hollow Rd, Melville, NY 11747

CHAIN-OF-CUSTODY Analytical Request Document

Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

LAB USE ONLY- Affix Workorder/Login Label Here

WO#: 70337400

Company Name: CITI BOCES
Street Address: 179 County Rte 64, Mexico, NY 13114

Contact/Report To: Terry Stoddard
Phone #: (315) 963-4251
E-Mail: tstoddard@ciboces.org
Cc E-Mail:

Customer Project #: CITI BOCES
Project Name:

Invoice To: Terry Stoddard
Invoice E-Mail: tstoddard@ciboces.org

Site Collection Info/Facility ID (as applicable):

Purchase Order # (if applicable):
Quote #:

Time Zone Collected: [] AK [] PT [] MT [] CT [X] ET
Data Deliverables:
[] Level II [] Level III [] Level IV
[] EQUIS
[] Other
Rush (Pre-approval required):
[] 2 Day [] 3 day [] 5 day [] Other
Date Results Requested: Standard 10 business day
Field Filtered (if applicable): [] Yes [] No
Analysis:

County / State origin of sample(s): New York

Regulatory Program (DW, RCRA, etc.) as applicable: NY Lead in School DW
DW PWSID # or WW Permit # as applicable:

* Matrix Codes (Insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk

Customer Sample ID	Matrix *	Comp / Grab	Collected		Composite End	Res. CL2	Number & Type of Containers		200.8 Drinking Water (Pb only)	Preservation non-conformance identified for sample.
			Date	Time	Date	Time	Plastic	Glass		
CITI 24	DW	G	2/16/2025	5:54			1		X	
CITI 27				6:11						
CITI 28				6:11						
CITI 35				6:05						
CITI 36				6:03						
CITI 37				6:03						
CITI 38				6:03						
CITI 40				6:05						
CITI 43				6:14						
CITI 44				6:14						

Customer Remarks / Special Conditions / Possible Hazards:
Lead

Collected By: Rich Paszkiewicz
Printed Name: Rich Paszkiewicz
Signature:

Relinquished by/Company: (Signature)
Relinquished by/Company: (Signature)
Relinquished by/Company: (Signature)
Relinquished by/Company: (Signature)

Date/Time: 2/16/25 13:55
Date/Time: 2/10
Date/Time: 2/11 12:50

Received by/Company: (Signature)
Received by/Company: (Signature)
Received by/Company: (Signature)
Received by/Company: (Signature)

Tracking Number:
Delivered by: [] In-Person [] Courier
[] FedEx [] UPS [] Other

Page: 1 of 7

CHAIN-OF-CUSTODY Analytical Request Document

Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

LAB USE ONLY- Affix Workorder/Login Label Here



Scan QR Code for instructions

Company Name:	CITI BOCES	Contact/Report To:	Terry Stoddard
Street Address:	179 County Rte 64, Mexico, NY 13114	Phone #:	(315) 963-4251
		E-Mail:	tstoddard@citi boces.org
		Cc E-Mail:	
Customer Project #:		Invoice To:	Terry Stoddard
(Project Name:	CITI BOCES	Invoice E-Mail:	tstoddard@citi boces.org
Site Collection Info/Facility ID (as applicable):		Purchase Order # (if applicable):	
		Quote #:	

Time Zone Collected: [] AK [] PT [] MT [] CT [] ET	County / State origin of sample(s):	New York
Data Deliverables:		
[] Level II [] Level III [] Level IV	Rush (Pre-approval required): [] 12 Day [] 3 day [] 5 day [] Other _____ Date Results Requested: Standard 10 business day	DW PWSID # or WW Permit # as applicable:
[] EQUIS		Field Filtered (if applicable): [] Yes [] No
[] Other _____		Analysis:
* Matrix Codes (Insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk		

Customer Sample ID	Matrix *	Comp / Grab	Collected (or Composite Start)		Composite End		Res. CL2	Number & Type Containers	
			Date	Time	Date	Time		Plastic	Glass
CITI 45	DW	G	2/6/2025	614				1	
CITI 48				639					
CITI 52				641					
CITI 57				644					
CITI 58				642					
CITI 60				606					
CITI 61				647					
CITI 64				653					
CITI 65				653					
CITI 68				648					

Customer Remarks / Special Conditions / Possible Hazards:	Collected By:
Lead	Printed Name: Rich Paszkiewicz Signature: <i>Rich Paszkiewicz</i>
Relinquished by/Company: (Signature)	Received by/Company: (Signature)
Relinquished by/Company: (Signature)	Received by/Company: (Signature)
Relinquished by/Company: (Signature)	Received by/Company: (Signature)
Relinquished by/Company: (Signature)	Received by/Company: (Signature)

[illegible]



Scan QR Code for instructions

Company Name:	CITI BOCES	Contact/Report To:	Terry Stoddard
Street Address:	179 County Rte 64, Mexico, NY 13114	Phone #:	(915) 963-4251
		E-Mail:	tstoddard@citi boces.org
		Cc E-Mail:	
Customer Project #:		Invoice To:	Terry Stoddard
Project Name:	CITI BOCES	Invoice E-Mail:	tstoddard@citi boces.org
Site Collection Info/Facility ID (as applicable):		Purchase Order # (if applicable):	
		Quote #:	

Time Zone Collected: [] AK [] PT [] MT [] CT [] ET	County / State origin of sample(s):	New York
Data Deliverables:		
([] Level II ([] Level III ([] Level IV	Rush (Pre-approval required): [] 12 Day [] 3 day [] 5 day [] Other _____	
([] EQUIS		
([] Other _____	Date Results Requested: Standard 10 business day	Field Filtered (if applicable): [] Yes [] No Analysis:
* Matrix Codes (Insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Biossay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk		

Customer Sample ID	Matrix *	Comp / Grab	Collected (or Composite Start)		Composite End		Res. CL2	Number & Type of Containers
			Date	Time	Date	Time		
CTT 103	DW	G	6/16/2025	529				1
CTT 104				529				
CTT 105				532				
CTT 110				530				
CTT 111				531				
CTT 112				529				
CTT 118				535				
CTT 120				545				
CTT 122				541				
CTT 123				541				

Customer Remarks / Special Conditions / Possible Hazards:	Lead	Collected By:
Relinquished by Company: (Signature)	Date/Time:	Printed Name: Rich Paszkiewicz
Relinquished by Company: (Signature)	Date/Time:	Signature:
Relinquished by Company: (Signature)	Date/Time:	Received by Company: (Signature)
Relinquished by Company: (Signature)	Date/Time:	Received by Company: (Signature)
Relinquished by Company: (Signature)	Date/Time:	Received by Company: (Signature)
Relinquished by Company: (Signature)	Date/Time:	Received by Company: (Signature)

Client:

Work ID:

Profile #:

COC Page

65

COC Page _____ of _____

Use Point Number Spreadsheet

Add SCLOGFD to first sample for field charge

Multiday Project

Add SCLOGFD to first sample for field charge

Project

Additional Comments

Sender Initials

Page 31 of 32

WO#: 70337400

PM: CMG Due Date: 02/25/25

CLIENT: CITI

[illegible]

	Glass	Plastic
VG9U	40mL unprns clear vial	BP4U 125mL unpreserved plastic
VG9C	40mL Ascorbic-HCl clear vial	BP3U 250mL unpreserved plastic
VG9H	40mL HCl clear vial	BP2U 500mL unpreserved plastic
VG9T	40mL Sulfuric clear vial	BP1U 1L unpreserved plastic
VG9V	40mL Na Thiosulfate vial	BP4N 125mL HNO3 plastic
DG9Y	40mL Citrate-Na Thiosulfate	BP3N 250mL HNO3 plastic
DG9P	40mL Amber vial - TSP	BP2N 500mL HNO3 plastic
DG9A	40mL Ascorbic/Malic Acid 40mL	BP2S 500mL H2SO4 plastic
DG6T	Na Thio 60mL Vial	BP2S 500mL H2SO4 plastic
DG6S	Ammonium Cl/CUSO4 40mL	BP3C NaOH 250mL bottle
DG5U	1L Unpres. Jar (cont Ed)	BP3T 250mL Trizma
WG9U	8oz clear soil jar	BP3S 250mL Ammonium Acetate
WG4Q	4oz clear soil jar	BP4R 250mL NaHSO4-NH4OH
		BP1Z 1L NaOH, Zn Acetate
		BP1N 1L HNO3 plastic
		BP1R 1L Thiosulfate Amber Bottle

	Misc.
SP5T	120mL Coliform Na Thio
R	Terracote Kit
WG2U	2oz Unpreserved Jar
WGFU	4oz Unpreserved Jar
WGKU	8oz Unpreserved Jar
WGDU	16oz Unpreserved Jar
ZPLC	Ziplock Bag
TEOL	Tedlar Bag
BGTH	1L HCL Clear Glass
GN	General
WP	Wipe
LLHG	Low Level Hg Bottles
BGIN	1L HN03 Clear Glass

IOC	
BP1U	1L unpreserved plastic
BP3N*	250mL HNO3 plastic
BP3C	250mL Sodium Hydroxide
AG2U	500mL unpres amber glass
BP3U	250mL unpreserved plastic

SOC	
VG8T	40mL Na Thio amber vial
DG9A	40mL Acetic acid / malic Acid vials
DG9V	Citrate/Na Thiosulfate 40mL
DG8T	Na Thiosulfate 60mL vial
DG6M	MonOAcetic/Na Thio 60mL
AG3U	250mL unpries amber glass
AG3T	Na Thiosulfate 250mL bottle
BP1B	Na Thiosulfate Amber bottle
AG1T	Na Thiosulfate 1L Amber
AG1A	525 x Chemical Erlen

Matrix	
WT	Water
SL	Solid
NAL	Non-aqueous Liquid
OL	Oil
WP	Wipe
DW	Drinking Water

Can also be a BP4N

Container Codes

WO#: 70337400

Client Name: Citi

Project: PM: CMG Due Date: 02/25/25

CLIENT: CITI

Courier: ☐ Fed Ex ☐ UPS ☐ USPS ☐ Client ☐ Commercial ☒ Pace ☐ Other

Tracking #:

Custody Seal on Cooler/Box Present: ☐ Yes ☒ No Seals intact: ☐ Yes ☒ No Temperature Blank Present: ☐ Yes ☒ No

Packing Material: ☐ Bubble Wrap ☐ Bubble Bags ☐ Ziploc ☒ None ☐ Other

Type of Ice: Wet Blue ☒ None

Thermometer Used: TH211 Correction Factor: 0 ☐ Samples on ice, cooling process has begun

Cooler Temperature(°C): 16.4 Cooler Temperature Corrected(°C): 16.4 Date/Time 5035A kits placed in freezer

Temp should be above freezing to 6.0°C

USDA Regulated Soil (☒ N/A, water sample)

Did samples originate in a quarantine zone within the United States: AL, AR, CA, FL, GA, ID, LA, MS, NC, NM, NY, OK, OR, SC, TN, TX, or VA (check map)? ☐ Yes ☒ No

Did samples originate from a foreign source including Hawaii and Puerto Rico)? ☐ Yes ☒ No

If Yes to either question, fill out a Regulated Soil Checklist (ENV-FRM-MELV-0076) and include with SCUR/COC paperwork.

Date and Initials of person examining contents: ABB 2/11/25

	COMMENTS:
Chain of Custody Present: ☒ Yes ☐ No	1.
Chain of Custody Filled Out: ☒ Yes ☐ No	2.
Chain of Custody Relinquished: ☒ Yes ☐ No	3.
Sampler Name & Signature on COC: ☒ Yes ☐ No ☐ N/A	4.
Samples Arrived within Hold Time: ☒ Yes ☐ No	5.
Short Hold Time Analysis (<72hr): ☐ Yes ☒ No	6.
Rush Turn Around Time Requested: ☐ Yes ☒ No	7.
Sufficient Volume: (Triple volume provided for MS/MSD) ☒ Yes ☐ No	8.
Correct Containers Used: ☒ Yes ☐ No	9.
-Pace Containers Used: ☒ Yes ☐ No	10.
Containers Intact: ☒ Yes ☐ No	10.
Filtered volume received for Dissolved tests ☐ Yes ☐ No ☒ N/A	11. Note: if sediment is visible in the dissolved container.
Sample Labels match COC: ☒ Yes ☐ No	12.
-Includes date/time/ID/Analysis Matrix: SL ☒ WT ☐ OIL ☐ OTHER	

Date and Initials of person checking preservation: ABB 2/11/25

All containers needing preservation have been ☒ Yes ☐ No ☐ N/A	13. ☐ HNO ₃ ☐ H ₂ SO ₄ ☐ NaOH ☐ HCl
pH paper Lot # 231224	Sample #
All containers needing preservation are found to be in compliance with method recommendation? (HNO ₃ , H ₂ SO ₄ , HCl, NaOH>9 Sulfide, ☒ Yes ☐ No ☐ N/A	
Exceptions: VOA, Coliform, TOC/DOC, Oil and Grease, DRO/8015 (water).	
Per Method, VOA pH is checked after analysis	
Samples checked for dechlorination: ☐ Yes ☐ No ☒ N/A	Initial when completed: Lot # of added preservative: Date/Time preservative added:
KI starch test strips Lot #	14.
Residual chlorine strips Lot #	Positive for Res. Chlorine? Y N
SM 4500 CN samples checked for sulf ☐ Yes ☐ No ☒ N/A	15.
Lead Acetate Strips Lot #	Positive for Sulfide? Y N
Headspace in ALK Bottle (>6mm): ☐ Yes ☐ No ☒ N/A	16.
Headspace in VOA Vials (>6mm): ☐ Yes ☐ No ☒ N/A	17.
Trip Blank Present: ☐ Yes ☐ No ☒ N/A	
Trip Blank Custody Seals Present ☐ Yes ☐ No ☒ N/A	

Client Notification/ Resolution:

Field Data Required? Y / N

Person Contacted:

Date/Time:

Comments/ Resolution:

* PM (Project Manager) review (which includes the SCUR) is documented electronically in LIMS.